



## Can I choose HPLC mobile phase percentages other than those used in the HSQ Spreadsheet – FAQ

**The HSQ Qualification Kit's™ gradient test calls for mobile phase compositions of 100/0% and 90/10% (for each of up to four solvent pairs).**

The user then inputs the absorbance values at each portion of the gradient into the HSQ Kit™ spreadsheet, and the software then automatically calculates the pump **accuracy**. In theory, percentages other than 90/10 could be used, such as 50/50. However, the default values shown in the Excel table are 90/10 and we would recommend the user should use these. The reason is that errors in the pump **accuracy** are most pronounced at the extreme percentages (e.g. 90/10) so any problems in the pump will be most apparent under these conditions.

Click [HERE](#) for PQ™ & HSQ Qualification Kits™ ordering information and pictures.

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)