



Why does the color of the Holmium Oxide solution for HPLC performance qualification look different than before - FAQ

The “apparent” color of the solutions that you see depends on the room lighting.

Under fluorescent lights, the solutions may appear bright pink. Under regular outdoor lighting they usually a dull yellow. This is normal as it is a wavelength calibration solution, so it absorbs different light differently.

There is no issue with the product, and it is still very usable.

Click [HERE](#) for Holmium Oxide, PQ & HSQ qualification kit ordering information.

Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com