



Why does the color of the Holmium Oxide solution for HPLC performance qualification look different than before - FAQ

The “apparent” color of the solutions that you see depends on the room lighting.

Under fluorescent lights, the solutions may appear bright pink. Under regular outdoor lighting they usually appear a dull yellow. This is normal as it is a wavelength calibration solution, so it absorbs different light differently.

There is no issue with the product, and it is still very usable.

Click [HERE](#) for Holmium Oxide, PQ & HSQ qualification kit ordering information.

Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 04-05-2024