

Can I use ion pairing reagents with the Cogent Diamond Hydride HPLC column - FAQ

Ion-pairing will not work with the Diamond Hydride column.

It separates compounds on the basis of polarity. Using an ion-pairing reagent reduces that polarity so the compound can be separated on a Reversed Phase column.



Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 07-22-2024