

What are the advantages and disadvantages of acid washing RSA vials – FAQ

Acid washing autosampler vials has consequences.

There is no reason to acid wash RSA vials. The main reason to wash ordinary vials is to reduce adsorption due to native silanols on the glass. Since RSA has virtually no silanols, this is not necessary.

Some of the consequences to acid washing include time and materials but also you run the risk of chloride and sodium residue left in the vials as well as other trace elements found in the reagents.



Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com