



What are the dead volumes of the MRQ and Fused Insert autosampler vials - FAQ

For both of these autosampler vials, the general residual volume is about 2 μL .

The dead volume is analogous to residual volume.

Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 07-03-2024