



Autoclaving glass inserts with plastic feet is not recommended – Tech Information

Autoclaving glass or plastic inserts with attached plastic springs is not supported. While the glass can withstand temperatures up to 1,000°F, the polyethylene spring will become deformed, may crack and will lose its functionality.

However, you can sterilize these inserts through ethylene oxide gas sterilization.



Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 05-10-2024