

Autoclaving glass inserts with plastic feet is not recommended - Tech Information

Autoclaving glass or plastic inserts with attached plastic springs is not supported. While the glass can withstand temperatures up to 1,000°F, the polyethylene spring will become deformed, may crack and will lose its functionality.

However, you can sterilize these inserts through ethylene oxide gas sterilization.

