



Can Cogent TYPE-C HPLC columns be used in supercritical fluid chromatography SFC – FAQ

As far as compatibility in SFC is concerned, the answer is yes.

These columns are fully compatible with the typical conditions used in an **SFC** application and the packing material and hardware will not be damaged by supercritical fluids used in **SFC**.

We are sorry but at this time, we cannot say whether they will be useful in your **SFC** application as we do not have **SFC** instrumentation in our laboratories and therefore cannot directly support their use. However, in terms of technical assistance, we are happy to provide advice and suggestions but not detailed applications at this time.



MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com