

## What do You suggest for an internal standard in glucosamine LCMS Analysis - FAQ

If you are using a  $^{13}\text{C}$ -glucosamine internal standard though, you may encounter problems with detection reproducibility. Matching the concentration reasonably well with the amount in the sample should give you no more than about 1% error, which should be tolerable unless you need extreme precision in the quantitation.

Alternatively, try using deuterated glucosamine as an internal standard instead. In the EICs, you can readily distinguish glucosamine from either the  $^{13}\text{C}$  isotope or the deuterated version. The principle is the same in either case.

The advantage of using the deuterated internal standard is that the natural abundance of deuterium is much lower than that of  $^{13}\text{C}$ .



*Glucosamine can be Retained on the Cogent Diamond Hydride™ HPLC Column.*

Click [HERE](#) for Cogent Diamond Hydride Ordering Information.

Printed from the Chrom Resource Center

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

Date: 06-16-2024