

In other retention modes such as ion exchange or HILIC, a high salt concentration may be required. However, these high levels can be detrimental to various aspects of an analysis. In LC-MS, high salt can suppress MS ionization. Even if the mobile doesn't use high salt concentrations, they can still elute off a column that had previously been used under such conditions.

With Cogent TYPE-C™ silica HPLC columns in **Aqueous Normal Phase (ANP)** mode, these high levels are not required; generally no more than 10mM should be used in any method, whether in LC-MS or regular HPLC. The salts can permanently alter the stationary phase behavior at these high levels and may be virtually impossible to completely remove, even with extensive washing protocols.



Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com