



Use Teflon bottles for A and B solvents in negative ion mode LCMS – Primer

You will want to have both your A and B solvents (*aqueous and organic*) in Teflon mobile phase reservoir bottles. The reason for using these bottles is that sodium present in borosilicate glass mobile phase reservoirs can lead to issues for anionic analytes in negative ion mode LCMS. These include phosphates and organic acids but can be others as well.

It is also advised that you consider using Reduced Surface **Activity** RSA™ glass autosampler vials for your samples to minimize sodium adducts, pH changes or basic compound adsorption.

Click [HERE](#) for ordering, technical information and pictures of RSA Autosampler Vials.

Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 05-19-2024