

How can I fix poor peak shape for phosphorylated compounds in aqueous normal phase ANP methods - FAQ

We have a number of strategies for fixing problems with peak shape for phosphorylated compounds:

- Ensure the column is not overloaded. We recommend an injection volume of 1 μ L or less with aqueous normal phase ANP methods.
- Use Teflon bottles instead of borosilicate glass for your mobile phase reservoirs. Sodium from the glass can cause peak shape problems for anionic compounds.
- Use 5-10 microM EDTA or medronic acid in the mobile phase and sample diluent. This will help chelate metals that cause problems for phosphates.
- Try adding 0.01% ammonia to the sample diluent.
- Try using a Cogent, metal free, coated stainless steel column.



Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 07-22-2024