

Suggestions on how to separate capsaicinoids in hot sauce by HPLC - Tips & Suggestions

If you have a hot sauce extract containing nordihydrocapsaicin, capsaicin, dihydrocapsaicin, and homodihydrocapsaicin, the Cogent Bidentate C18 2.0^{TM} HPLC column is a good choice for sufficient separation when the compounds are analyzed by LC-MS; look for the $[M+H]^+$ ions.


The following gradient should be a good starting point:

Mobile Phase: A: DI H₂O / 0.1% formic acid (v/v)

Gradient: B: Methanol

time (min.)	%B
0	10
1	10
9	70
10	70
11	10

You may get EICs which look like this:



- 1. nordihydrocapsaicin, 294.2064 m/z [M+H]⁺
 - 2. capsaicin, 306.2064 m/z [M+H]⁺
- 3. dihydrocapsaicin, 308.2220 m/z [M+H]⁺



4. homodihydrocapsaicin, 322.2377 m/z [M+H]⁺

For peaks 1 and 2, you will be able to obtain separate peaks for each compound in the EICs so there will not be an issue of interference from co-elution.

In terms of sample preparation, try using a Soxhlet extraction procedure.

Click HERE for Cogent Bidentate C18 HPLC Column Ordering Information.



Printed from the Chrom Resource Center Copyright 2024, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451 tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com
Date: 09-19-2024