

To establish *linearity* of the proposed method, ten point calibration curves should be generated with appropriate concentrations of calibration standard solutions. The calibration range can be for example: 10-100 microg/mL.

The linearity should be evaluated by the least squares regression method using unweight data.



Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com