
The effect of ionic strength of the mobile phase on retention for Cogent HPLC columns in Aqueous Normal Phase (ANP) HPLC Analyses.

Ionic strength has a smaller effect than in HILIC because it appears that the ANP mechanism is some type of competitive adsorption in the adsorbed layer of acetonitrile on the surface.

Ionic strength would be expected to have some impact on this process.

