

## If an analyte is only soluble at basic pH can you use a basic mobile phase for my method – FAQ

**Use of a strongly basic mobile phase is usually not recommended for any silica based HPLC column** because the underlying silica of the stationary phase is prone to dissolution at pH above 8 when the solution has a strong ionic strength. This may seem problematic for some analytes that you want to use in **Aqueous Normal Phase ANP** or Reversed Phase HPLC.

Folic acid for example, is not appreciably soluble in neutral to acidic conditions.

However, you can use a basic diluent for the sample only and still use a mild or acidic pH mobile phase ANP. The small volume of the sample plug will not cause significant dissolution of the stationary phase and will allow for accurate quantitation of the **analyte**.



Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

Date: 09-26-2024