



Since Cogent TYPE-C columns can be used in Reversed Phase or Normal Phase what do you suggest to switch these solvent systems which are not miscible - FAQ

Suggested Procedure:

A - Changing from Reversed Phase to Normal Phase HPLC; pump 100% methanol for 15 minutes at 1 mL / minute flow rate, followed by 15 minutes 100% methylene chloride. The column is ready to be equilibrated with mobile phase for NP HPLC.

B - Changing from Normal Phase to Reversed Phase HPLC; pump 100% methylene chloride for 15 minutes at 1 mL / minute flow rate, followed by 15 minutes 100% methanol. The column is ready to be equilibrated with mobile phase for RP-HPLC.

A simple 30 minute procedure allows switching from one mode to another.



MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com