

Using two HPLC columns together in a tandem separation for polar and non polar retention - Tips & Suggestions

Use two columns and two modes of HPLC together for excellent results.

If you connect the Cogent Bidentate C18™ & Cogent Diamond Hydride™ HPLC columns in parallel, but use the same mobile phase program on both columns, it will be difficult to have both Reversed Phase and **Aqueous Normal Phase ANP** HPLC modes operating simultaneously, therefore you will not get optimal results. A gradient starting at high organic may work well for the Cogent Diamond Hydride™ HPLC column but only very hydrophobic compounds will retain on the Cogent Bidentate C18™ HPLC column.

However, the attached poster, linked below, describes a more sophisticated system where retention of both hydrophobic and hydrophilic analytes can be retained using both columns. The set-up and procedure are described in detail.

Metabolomics studies involve not only known and predicted metabolites but also of a wide variety of ones that had not previously been predicted to exist. Due to the significant complexity of any metabolome, both the Bidentate C18™ and Diamond Hydride™ together give a much more complete picture than when used individually.

NOTE: You need appropriate instrumentation to get the best results. Both columns are available together:

[Metabolomics Kit Product Page](#)



Attachment:

University of Melbourne Poster [Download File](#)

Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 07-03-2024