



Dantrolene Sodium Analyzed with HPLC with UV – AppNote

Dantrolene Sodium has several amine groups that can interact with lone-silanols causing problematic peak tailing when analyzed with conventional HPLC columns. By utilizing a Bidentate C18™ Column with its Silica Hydride™ surface, challenging compounds like Dantrolene can be readily retained with symmetrical peak shape.



Dantrolene sodium

Peak:

Dantrolene Sodium

Method Conditions:

Column: Cogent Bidentate C18™, 4µm, 100Å

Catalog No.: 40018-75P

Dimensions: 4.6 x 75 mm

Mobile Phase: 60% DI Water / 40% Acetonitrile / 0.1% Formic Acid (v/v)

Injection Volume: 2µL

Flow Rate: 1.0ml / min

Detection: 225 nm

Samples: 0.1 mg/mL Dantrolene Sodium in 50:50 Acetonitrile: DI H2O

Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

Note: Essential to muscle contraction are Ryanodine Receptors that regulate the release of Calcium from the Sarcoplasmic Reticulum of muscle cells. Dantrolene Sodium is a postsynaptic muscle relaxant that lessens the "excitation-contraction" coupling response in these cells. It achieves this by inhibiting Calcium binding to Ryanodine

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Receptor 1 and decreasing intracellular Calcium concentrations.



Attachment

A384 Dantrolene AppNote pdf 0.1 Mb [Download File](#)

Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com