

Can agents to passivate columns be used on Cogent TYPE-C HPLC columns? – FAQ

Yes, they can be used with Cogent TYPE-C™ HPLC Columns.

Sometimes an acidic or basic condition is required to improve peak shape or sensitivity. Often when negatively charged compounds are analyzed with mass spectrometry, buffers are used to enhance sensitivity. Poor peak shape is caused by metals, particularly iron, and components leaching from an HPLC system or the column hardware.

Medronic acid, a methylene-substituted bisphosphonate, for example, can be used to mitigate these peak shape problems and is compatible with the Cogent TYPE-C™ columns.



Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 10-13-2024