



Can I use chlorinated or fluorinated solvents with PEEK frits and columns for HPLC - FAQ

Yes!

PEEK is compatible with most halogenated organic solvents that are used in HPLC.



Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 05-20-2024