



## Can I use chlorinated or fluorinated solvents with PEEK frits and columns for HPLC – FAQ

Yes!

PEEK is compatible with most halogenated organic solvents that are used in HPLC.



Printed from the Chrom Resource Center

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

Date: 05-21-2024