



Are the MicroSolv mobile phase bottles Type 1, Class A glass - FAQ

Yes, they are.

Both the safety-coated and non-coated MicroSolv mobile phase bottles are crafted from high-quality borosilicate 3.3 glass, a type of Type 1, Class A borosilicate glass. This material offers excellent chemical resistance, making these bottles suitable for a wide range of laboratory applications.

- Click [HERE](#) for ordering information for safety coated bottles.
- Click [HERE](#) for ordering information for non - coated bottles.

Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com