



Can you use urea buffers with PEEK tubing – FAQ

Long term use of urea buffers are not indicated for use with PEEK as it can weaken the tubing material and therefore should be avoided in HPLC when possible.

For short term use, you should flush the tubing thoroughly after each run.

Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 05-19-2024