


## Fluorouracil Assay Analyzed with HPLC - AppNote

## A Reliable Method for a Chemotherapeutic Drug

This Fluorouracil Assay is easily performed and demonstrates Run to Run consistency and Precision with Retention Time RSD values below 0.1%. This Method shows reliability for analysis of this Pyrimidine Analog that is an antineoplastic anti-metabolite.



Peak:

Fluorouracil

## **Method Conditions:**

Column: Cogent RP C18™, 5um, 100Å

Catalog No.: 68518-25P

Dimensions: 4.6mm x 250mm

Mobile Phase: (5:95) Methanol / Buffer

Buffer: 0.73g / L of Monobasic Sodium Phosphate and 1.4g / L of Dibasic Sodium Phosphate in Water.

Injection vol.: 20µL

Flow rate: 1.0mL / minute Detection: UV @ 254nm

Sample Preparation: Fluorouracil 1.0mg / mL in DI Water

**Note:** Fluorouracil is a Pyrimidine Analog used to treat Basal Cell Carcinomas, and as an injection in palliative Cancer treatment.



This Method was developed by and is presented courtesy of ARL- Eutech Scientific Services, Inc.



Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451 tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 07-27-2024