



# GRL-0617, a benzamide anti-Covid molecule and potent non-covalent inhibitor

This simple gradient analyzing GRL-0617 is easily performed, and the 5 injections overlay below demonstrates run to run consistency and great peak symmetry. The peak efficiency is very good, peak tailing is minimal, and repeatability is easily met with RSD values being less than 1.0%.





# Peak:

GRL0617

## **Method Conditions:**

Column: Cogent RP Phenyl Hexyl™, 5um, 100Å

Catalog No.: 68539-15P

**Dimensions:** 4.6mm x 150mm

Mobile phase:

A: DI water and 0.1% formic acid
B: Acetonitrile and 0.1% formic acid

#### **Gradient:**

time (minutes)	%B
0	25
1	25

Printed from the Chrom Resource Center Copyright 2024, All Rights Apply

### **MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451 tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com



5	85
6	85
8	25
10	25

Flow rate: 1.0 mL / minute Injection volume:  $1 \mu \text{L}$  Detection: UV @ 224nm

**Sample Preparation:** 1.0 mg/mL of GRL-0617 in methanol.

**%RSD:** <0.5% **t₀:** 2.0 minutes

**K':** 2.075

**Note 1:** GRL0617 is an inhibitor of SARS-CoV and SARS-CoV-2 papain-like proteases (PL-pro). It is selective for the viral proteases over human proteases. GRL0617 blocks both PL-pro-mediated actions towards host biological processes and virus replication.

Note 2: Capacity is determined using the following equation:  $k = (t_R - t_0)/t_0$ 

- $t_R$  = Retention Time of an Analyte Peak
- $t_0$  = Retention Time of non-Retained Peak



Printed from the Chrom Resource Center

Copyright 2024, All Rights Apply

**MicroSolv Technology Corporation** 

9158 Industrial Blvd. NE, Leland, NC 28451 tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com