


Alternate L1 column used for a similar Hydrocortisone Assay with HPLC - AppNote


Alternate L1 column used for a similar USP assay method of Hydrocortisone

This Similar USP Hydrocortisone Assay is easily performed and the five injection overlay below demonstrates Run to Run consistency and Precision. The Peak Efficiency is very good and Peak Tailing guidelines are easily met with RSD values of 0.5%.



This demonstrates a good alternate Column for your Hydrocortisone USP Method.

Peak:



Hydrocortisone

Method Conditions:

Column: Cogent MicroBee C18™, 5um, 125Å

MICROSOLV

Catalog No.: 59305-03

Dimensions: 3.9mm x 300mm

Mobile Phase: 25:75 Acetonitrile / DI Water

Injection vol.: 15µL

Flow rate: 1.0mL / minute

Detection: UV @ 254nm

Sample Preparation: Hydrocortisone 50µg / mL in Methanol

Note 1: *Hydrocortisone, a Corticosteroid, and is similar to a natural hormone produced by your Adrenal Glands. Hydrocortisone is used to treat immune, inflammatory, and neoplastic conditions.*

Note 2: *The Cogent MicroBee C18™ performs very similarly to the Waters® µBondapak C18®.*



Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 07-22-2024