



Is there residue of phosphoric acid in PEEK products – FAQ

Yes, there is a possibility of phosphoric acid residue in every type and brand of PEEK material (tubing and other PEEK manufactured items).

This is a known side effect from the manufacturing process, and it should be noted only on a few occasions PEEK parts may contain only a very small amount of phosphoric acid residue, which is washed away rapidly and will not last long, if at all.

This small amount of phosphoric acid would most likely only *briefly* effect critical methods and not large scale/prep chromatography or chemistry.

Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 05-17-2024