


## Triton X-100 Assay by HPLC- AppNote

### A Sensitive Assay for Detection of a Non-ionic Surfactant

Triton X-100 is used in the sample purification steps of various materials as well as for cleaning apparatus in pharmaceutical manufacturing. Due to the toxic nature of Triton X-100, a sensitive method for its detection and quantification is very important to identify any residual amounts.

This simple Isocratic Method may be used to verify residual amounts of this Surfactant have been removed from final material or from vital manufacturing equipment.



**Peak:**

Triton X-100

### Method Conditions:

**Column:** Cogent Bidentate C18™, 4µm, 100Å

**Catalog No.:** 40018-75P

**Dimensions:** 4.6 x 75mm

**Mobile Phase:** 80:20 Acetonitrile / DI Water with 0.1% Formic Acid

**Injection vol.:** 5µL

**Flow rate:** 1.0mL / minute

**Detection:** UV @ 234nm

**Sample Preparation:** Triton X-100, 50ug / mL in DI Water

**Note:** Triton X-100 is one of the most widely used nonionic surfactants for lysing cells to extract protein and other cellular organelles. Triton X-100 ( $C_{14}H_{22}O(C_2H_4O)_n$ ) has a hydrophilic polyethylene oxide chain and an aromatic hydrocarbon lipophilic or hydrophobic group. The hydrocarbon group is a 4-phenyl group.

This Method was developed by and is presented courtesy of **ARL- Eutech Scientific Services, Inc.**



Printed from the Chrom Resource Center

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

Date: 05-14-2024