

Which column coupler do you recommend for pairing HPLC columns with guard columns- Tech Information

The ARE-Applied Research™ brand of column couplers are used to connect the flow of two HPLC columns or a guard column and an analytical column directly in series.

Below is our recommended selection chart to match your flow rate and a suggested union-column coupler (ID). The ID of the coupler should be determined based on your application, where the faster the flow, the larger the bore (ID) is needed.

Flow rate <i>(milliliters / minute)</i>	Through Hole <i>(mm)</i>	Through Hole <i>(inch)</i>	Column Coupler <i>Cat. No.</i>
up to 0.5	0.13	0.005	47450-01
≤1.0	0.18	0.007	47450-02
1.0 - 2.0	0.25	0.010	47450-03
≥2.0	0.50	0.020	47450-04

ARE-APPLIED-RESEARCH™

Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 07-22-2024