

What parts will I need for Luer fittings and a flash chromatography column – FAQ

I'm trying to hook up a flash chromatography column (female Luer inlet, male Luer outlet) to couple with my old HPLC system. The HPLC system has the typical 1/16" stainless steel tubing and 10-32 threaded fittings. What would I need to connect them?

To connect the column to the HPLC system, you will need the following:

- A. [47451-42](#) Adapter: 10-32 female to male luer connection (for inlet)
- B. 53.9.200124 Adapter: PP, female Luer to 10-32 male (for outlet)



Printed from the Chrom Resource Center

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Date: 05-16-2024