



Injection Technique for Aqueous Samples using Cogent TYPE-C Columns in ANP - Tips & Suggestions

Some samples can be in 100% Water Diluents when using Aqueous Normal Phase (ANP) HPLC.

Although it is preferable to use a high organic content diluent for your sample in ANP, we have observed excellent peak shapes using entirely aqueous injections in many applications.

This [application note](#) featuring the Cogent Diamond Hydride™ Column is an example of analysis of folic acid in standards and juice samples in ANP. A 100% water-based diluent was used and the peak shape was both sharp and symmetrical.

[Click HERE for Cogent Diamond Hydride HPLC Column Ordering Information.](#)



Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com