



Why do you have different bonded phases with the Cogent TYPE-C columns - FAQ

INTERNAL ONLY: The degree of analyte retention in Reversed Phase (*RP*) and Aqueous Normal Phase (*ANP*) on Cogent TYPE-C columns depends on the degree and type of modification to the stationary phase.

For no or minimal modification the ANP mode is stronger than RP mode.

As the degree of modification becomes greater, i.e. larger groups and/or higher surface coverage, the RP properties increase.



Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com