



Gemcitabine Assay Analyzed with HPLC – AppNote

A Reliable Method for a Chemotherapeutic Drug

This Gemcitabine Assay is easily performed and demonstrates Run to Run consistency and Precision with Retention Time RSD values below 0.4%. This Method shows consistency and reliability for analysis of this Nucleoside Chemotherapeutic.





Method Conditions:

Column: Cogent RP C8[™], 5um, 150Å Catalog No.: <u>68508-25P</u> Dimensions: 4.6mm x 250mm Mobile Phase: 13.8g of Monobasic Sodium Phosphate and 2.5ml of Phosphoric Acid in 1L of Water. Injection vol.: 20µL Flow rate: 1.2mL / minute Detection: UV @ 275nm Sample Preparation: Gemcitabine 1.0mg / mL in DI Water of certain types of ovarian cancer, non-small cell lung carcinoma, metastatic breast cancer, and as a single agent for pancreatic cancer.

This Method was developed by and is presented courtesy of <u>ARL- Eutech Scientific Services, Inc.</u>



Printed from the Chrom Resource Center Copyright 2025, All Rights Apply **MicroSolv Technology Corporation** 9158 Industrial Blvd. NE, Leland, NC 28451 Tel: (732) 380-8900 Fax: (910) 769-9435 Email: customers@mtc-usa.com Website: www.mtc-usa.com