

## Can amber or clear polypropylene vials be sterilized with autoclaving or ethylene oxide - FAQ

Polypropylene autosampler vials [9502S-PP-Clear](#) should not be autoclaved because they may warp during heating. However, you can use ethylene oxide gas to sterilize the vials if you do not exceed 120°C.

Click [HERE](#) for more information about polypropylene vials and pictures

Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)