

Can I use a laser to make a detection window for the Controlled Flow and Zero Flow capillaries - FAQ

No, these capillaries have a special, covalently bonded coating on the inside capillary wall (linear polyacrylamide) that may be damaged using a laser or heat.

The outer polyimide coating can be removed by a [specific protocol](#) for this purpose. This procedure uses fuming sulfuric acid to chemically remove the outer polyimide coating.



Click [HERE](#) for ordering information and pictures.

Printed from the Chrom Resource Center
Copyright 2025, All Rights Apply
MicroSolv Technology Corporation
9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900
Fax: (910) 769-9435
Email: customers@mtc-usa.com
Website: www.mtc-usa.com