



## How resistant are the ultra fine point markers to smearing - FAQ

The markers were tested on standard glass autosampler vials and the writing was found to be completely smear-resistant when using a dry cloth, after waiting a short while ( about 30 seconds) for the ink to dry.

If you wish to remove the writing however, this can be accomplished with a solvent such as methanol, acetone, or isopropanol.

Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)