

Can the BASIK brand filter degasser be used to de-gas carbon dioxide from a solvent - FAQ

Under normal conditions, yes you can use the filter degasser device to remove carbon dioxide CO₂.

In terms of the performance of this unit, the intended scope of the product is for removal of particulate matter and gas bubbles that can appear in HPLC solvents. The filter membrane that is included for particulates is compatible with most HPLC solvents and the vacuum that is needed for this unit to work, will remove dissolved gases such as oxygen & carbon dioxide from the HPLC solvent.

It is important to note that exposure to the atmosphere will re-dissolve gases in the solvents. Not all solutions or solvents will be compatible with this unit.

Click [HERE](#) for BASIK™ brand glass filter degasser device ordering information and pictures.

Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com