

What is the best way to attach the inlet of the knitted reactor coil KRC tubing to the outlet of a tee with a PEEK finger tight fitting that will hold high back pressure - FAQ

Confirm that there is not an impediment to reasonable flow through the tubing which could increase the back pressure. We would then recommend using our items [47320-30](#) and [47320-34](#) which are high pressure PEEK HPLC tubing connectors or male nuts and ferrules.

Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com