



Will using the Cogent Column Filters change my current HPLC results - FAQ

No, the pre-column, Cogent Column Filters will have no effect on your chromatography if your analytes are compatible with PEEK and stainless steel.

It will simply protect the column from dust, non-dissolved salts, pump seal particles or dirt that passes from the HPLC pump into a column.

[Click here to view more information about pre-column filters for HPLC](#)



Printed from the Chrom Resource Center
Copyright 2025, All Rights Apply
MicroSolv Technology Corporation
9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900
Fax: (910) 769-9435
Email: customers@mtc-usa.com
Website: www.mtc-usa.com