

What is the shelf life and expiration date of the solutions in the HSQ Kit - FAQ

Shelf life and stability data for the HSQ Kit™ has been studied and collected for 24-month periods many times for many years with no observed instability. Therefore, the expiration date is set for two years after the date of manufacture based on the stability testing done and to satisfy most customer internal policies.

Kits available for purchase will have at least six months left until the expiration date which should be a long enough period of time to properly use the kits..

Click [HERE](#) for PQ & HSQ Kits™ for HPLC ordering information and pictures.

Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com