



What are some common sources of variance during sample preparation in an analytical application - FAQ

Each step in a sample preparation procedure can introduce error to your analytical results. Therefore, it is important to know what equipment or techniques contribute to the error and how significant each one is.

For example, weighing is extremely common in sample preparation and typically has a very low variance due to the accuracy of analytical scale balances. Indeed, gravimetric assays are among the most accurate procedures available.

However, some other factors should be adequately considered in contributing to the total analytical variance. Volumetric flasks have inherent tolerances that can be considered for optimal method accuracy. For example, you could make a solution of a given concentration using a 250 mL or a 25 mL volumetric flask, but the former has a lower tolerance.



Printed from the Chrom Resource Center
Copyright 2025, All Rights Apply
MicroSolv Technology Corporation
9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900
Fax: (910) 769-9435
Email: customers@mtc-usa.com
Website: www.mtc-usa.com